



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

The Meter Shop
6934 Signat Drive, Houston Texas 77041

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Electrical and Mechanical Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Initial Accreditation Date:

February 23, 2016

Issue Date:

May 27, 2020

Expiration Date:

May 31, 2022

Accreditation No.:

83790

Certificate No.:

L20-314

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signal Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	
Equipment to Output DC Voltage ^F	2 mV to 220 mV	9.5 μ V/V + 0.4 μ V	Fluke 5520A GIDEP/MET TRACK	
	0.22 V to 2.2 V	6 μ V/V + 0.7 μ V		
	2.2 V to 11 V	4.8 μ V/V + 2.5 μ V		
	11 V to 22 V	6.8 μ V/V + 4 μ V		
	22 V to 220 V	5.9 μ V/V + 40 μ V		
	220 V to 1 100 V	8.7 μ V/V + 0.4 mV		
Equipment to Measure DC Voltage ^{FO}	2 mV to 200 mV	9.4 μ V/V + 0.1 μ V	Fluke 8508A GIDEP/MET TRACK	
	200 mV to 2 V	4.2 μ V/V + 0.4 μ V		
	2 V to 20 V	4.7 μ V/V + 4 μ V		
	20 V to 200 V	6.4 μ V/V + 40 μ V		
	200 V to 1 000 V	6.4 μ V/V + 0.5 mV		
	1 kV to 40 kV	0.11 %		Direct measurements using Fluke 80K-40 Probe/Fluke 8508A
Equipment to Output DC Current ^{FO}	2 μ A to 22 μ A	52 μ A/A + 6 mA	Fluke 5520A GIDEP/MET TRACK	
	220 μ A to 2.2 mA	45 μ A/A + 7 mA		
	2.2 mA to 22 mA	46 μ A/A + 40 mA		
	22 mA to 220 mA	0.39 mA/A + 0.5 mA		
	220 mA to 2.2 A	2 mA/A + 0.5 mA		
	2.2 A to 20 A	0.4 mA/A + 0.75 mA		
	20 A Turns to 150 A Turns	0.5 % + 0.15 A		Fluke 5520A w/coil GIDEP/MET TRACK
	150 A Turns to 1 000 A Turns	0.52 % + 0.55 A		
Equipment to Measure DC Current ^{FO}	20 μ A to 200 μ A	67 μ A/A + 0.4 mA	Fluke 8508A GIDEP/MET TRACK	
	200 μ A to 2 mA	67 μ A/A + 4 mA		
	2 mA to 20 mA	68 μ A/A + 40 mA		
	2 mA to 200 mA	86 μ A/A + 0.8 μ A		
	200 mA to 2 000 mA	58 μ A/A + 16 μ A		
	2 A to 20 A	58 mA/A + 0.4 mA		
Equipment to Output DC Resistance ^{FO}	0.1 Ω to 1 Ω	13 $\mu\Omega/\Omega$ + 40 $\mu\Omega$	Fluke 5520A GIDEP/MET TRACK	
	1 Ω to 1.9 Ω	14 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	1.9 Ω to 10 Ω	30 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	10 Ω to 19 Ω	43 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	19 Ω to 100 Ω	14 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	100 Ω to 190 Ω	18 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	190 Ω to 1 000 Ω	13 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	1 k Ω to 1.9 k Ω	14 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	1.9 k Ω to 10 k Ω	13 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	10 k Ω to 19 k Ω	14 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	19 k Ω to 100 k Ω	15 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	100 k Ω to 190 k Ω	14 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	190 k Ω to 1 M Ω	25 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	1 M Ω to 1.9 M Ω	28 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	1.9 M Ω to 10 M Ω	48 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
	10 M Ω to 19 M Ω	62 $\mu\Omega/\Omega$ + 40 $\mu\Omega$		
19 M Ω to 100 M Ω	14 $\mu\Omega/\Omega$ + 40 $\mu\Omega$			



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signat Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Resistance ^{FO}	0 Ω to 2 Ω	25 $\mu\Omega/\Omega$ + 4 $\mu\Omega$	Fluke 8508A GIDEP/MET TRACK
	2 Ω to 20 Ω	12 $\mu\Omega/\Omega$ + 14 $\mu\Omega$	
	20 Ω to 200 Ω	9.4 $\mu\Omega/\Omega$ + 50 $\mu\Omega$	
	200 Ω to 2 000 Ω	9.4 $\mu\Omega/\Omega$ + 0.5 m Ω	
	2 k Ω to 20 k Ω	9.4 $\mu\Omega/\Omega$ + 5 m Ω	
	20 k Ω to 200 k Ω	9.5 $\mu\Omega/\Omega$ + 50 m Ω	
	0.2 M Ω to 2 M Ω	14 $\mu\Omega/\Omega$ + 1 Ω	
	2 M Ω to 20 M Ω	57 $\mu\Omega/\Omega$ + 0.1 k Ω	
	20 M Ω to 200 M Ω	92 $\mu\Omega/\Omega$ + 10 k Ω	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type B ^{FO}	600 $^{\circ}\text{C}$ to 800 $^{\circ}\text{C}$	0.35 $^{\circ}\text{C}$	Fluke 5520A Electrical Simulation of Thermocouple Output GIDEP/MET TRACK
	800 $^{\circ}\text{C}$ to 1 000 $^{\circ}\text{C}$	0.27 $^{\circ}\text{C}$	
	1 000 $^{\circ}\text{C}$ to 1 550 $^{\circ}\text{C}$	0.24 $^{\circ}\text{C}$	
	1 550 $^{\circ}\text{C}$ to 1 820 $^{\circ}\text{C}$	0.26 $^{\circ}\text{C}$	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type E ^{FO}	-250 $^{\circ}\text{C}$ to -100 $^{\circ}\text{C}$	0.39 $^{\circ}\text{C}$	
	-100 $^{\circ}\text{C}$ to -25 $^{\circ}\text{C}$	0.14 $^{\circ}\text{C}$	
	-25 $^{\circ}\text{C}$ to 350 $^{\circ}\text{C}$	0.12 $^{\circ}\text{C}$	
	350 $^{\circ}\text{C}$ to 650 $^{\circ}\text{C}$	0.14 $^{\circ}\text{C}$	
	650 $^{\circ}\text{C}$ to 1 000 $^{\circ}\text{C}$	0.17 $^{\circ}\text{C}$	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type J ^{FO}	-210 $^{\circ}\text{C}$ to -100 $^{\circ}\text{C}$	0.22 $^{\circ}\text{C}$	
	-100 $^{\circ}\text{C}$ to -30 $^{\circ}\text{C}$	0.14 $^{\circ}\text{C}$	
	-30 $^{\circ}\text{C}$ to 150 $^{\circ}\text{C}$	0.12 $^{\circ}\text{C}$	
	150 $^{\circ}\text{C}$ to 760 $^{\circ}\text{C}$	0.14 $^{\circ}\text{C}$	
	760 $^{\circ}\text{C}$ to 1 200 $^{\circ}\text{C}$	0.19 $^{\circ}\text{C}$	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type K ^{FO}	-200 $^{\circ}\text{C}$ to -100 $^{\circ}\text{C}$	0.26 $^{\circ}\text{C}$	
	-100 $^{\circ}\text{C}$ to -25 $^{\circ}\text{C}$	0.15 $^{\circ}\text{C}$	
	-25 $^{\circ}\text{C}$ to 120 $^{\circ}\text{C}$	0.14 $^{\circ}\text{C}$	
	120 $^{\circ}\text{C}$ to 1 000 $^{\circ}\text{C}$	0.21 $^{\circ}\text{C}$	
	1 000 $^{\circ}\text{C}$ to 1 372 $^{\circ}\text{C}$	0.33 $^{\circ}\text{C}$	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type N ^{FO}	-200 $^{\circ}\text{C}$ to -100 $^{\circ}\text{C}$	0.31 $^{\circ}\text{C}$	
	-100 $^{\circ}\text{C}$ to -25 $^{\circ}\text{C}$	0.18 $^{\circ}\text{C}$	
	-25 $^{\circ}\text{C}$ to 120 $^{\circ}\text{C}$	0.16 $^{\circ}\text{C}$	
	120 $^{\circ}\text{C}$ to 410 $^{\circ}\text{C}$	0.15 $^{\circ}\text{C}$	
	410 $^{\circ}\text{C}$ to 1 300 $^{\circ}\text{C}$	0.22 $^{\circ}\text{C}$	



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signat Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Electrical Calibration of Thermocouple Indicators ^{FO}			Fluke 5520A Electrical Simulation of Thermocouple Output GIDEP/MET TRACK
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type R ^{FO}	0 °C to 250 °C	0.45 °C	
	350 °C to 400 °C	0.28 °C	
	400 °C to 1 000 °C	0.26 °C	
	1 000 °C to 1 767 °C	0.31 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type S ^{FO}	0 °C to 250 °C	0.37 °C	
	250 °C to 1 000 °C	0.28 °C	
	1 000 °C to 1 400 °C	0.29 °C	
	1 400 °C to 1 767 °C	0.36 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type T ^{FO}	-250 °C to -150 °C	0.49 °C	
	-150 °C to 0 °C	0.19 °C	
	0 °C to 120 °C	0.14 °C	
	120 °C to 400 °C	0.15 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type U ^{FO}	-200 °C to 0 °C	0.44 °C	
	0 °C to 600 °C	0.22 °C	
50 Ω Load ^{FO}	1 mV to 24.999 mV	1.3 % of output + 40 μ V	
	25 mV to 2.199 9 V	0.23 % of output + 40 μ V	
	2.2 V to 130 V	2.1 % of output + 40 μ V	
1 M Ω Load ^{FO}	1 mV to 24.999 mV	0.12 % of output + 40 μ V	
	25 mV to 109.99 mV	0.09 % of output + 40 μ V	
	110 mV to 2.199 9 V	0.31 % of output + 40 μ V	
	2.2 V to 10.999 V	0.27 % of output + 40 μ V	
	11 V to 130 V	2 % of output + 40 μ V	
Level Sine Wave ^{FO}			
Amplitude (50 kHz Reference) FO	50 kHz	1.6 % + 300 μ V	
	50 kHz to 100 MHz	2.8 % + 300 μ V	
	100 MHz to 300 MHz	3.2 % + 300 μ V	
	300 MHz to 600 MHz	4.7 % + 300 μ V	
	600 MHz to 1 100 MHz	5.7 % + 300 μ V	



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signat Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Flatness 50 kHz Reference ^{FO}	50 kHz to 100 MHz	1.4 % + 100 μ Hz	Fluke 5520A/SC600 GIDEP/MET TRACK
	100 MHz to 300 MHz	1.6 % + 100 μ Hz	
	300 MHz to 600 MHz	3.2 % + 100 μ Hz	
	600 MHz to 1 100 MHz	3.9 % + 100 μ Hz	
Time Markers: Source and Period into a 50 Ω Load ^{FO}	5 ms to 50 ms	160 μ s + <i>t</i> parts in 106 s	
	20 ms to 2 ms	64 μ s + <i>t</i> parts in 106 s	
Equipment to Measure Amplitude At the listed Frequencies ^{FO}			
1 M Ω	1.8 mV(pk – pk) to 55 V (pk – pk)	2.4 % + 100 μ V	Fluke 5520A GIDEP/MET TRACK
50 Ω	1.8 mV(pk – pk) to 2.5 V (pk – pk)	2.4 % + 100 μ V	
Frequency ^{FO}	10 kHz to 100 kHz	34 parts in 106 Hz + 15 mHz	
Equipment to Measure AC Voltage Output At the listed Frequencies ^{FO}			
45 kHz to 1 kHz	3.3 V to 32.999 9 V	0.054 % + 650 μ V	
1 kHz 10 kHz	3.3 V to 32.999 9 V	0.026 % + 600 μ V	
10 kHz to 20 kHz	3.3 V to 32.999 9 V	0.11 % + 600 μ V	
20 kHz to 50 kHz	3.3 V to 32.999 9 V	0.16 % + 600 μ V	
50 kHz to 100 kHz	3.3 V to 32.999 9 V	0.17 % + 1.6 mV	
Equipment to Measure AC Voltage Output At the listed Frequencies ^{FO}			
45 kHz to 1 kHz	33 V to 329.999 V	0.024 % + 2 mV	
1 kHz to 10 kHz	33 V to 329.999 V	0.24 % + 6 mV	
10 kHz to 20 kHz	33 V to 329.999 V	0.25 % + 6 mV	
20 kHz to 50 kHz	33 V to 329.999 V	0.4 % + 6 mV	
50 kHz to 100 kHz	33 V to 329.999 V	0.43 % + 50 mV	
Equipment to Measure AC Voltage Output At the listed Frequencies ^{FO}			
45 kHz to 1 kHz	330 V to 1 000 V	0.41 % + 10 mV	
1 kHz to 5 kHz	330 V to 1 000 V	0.41 % + 10 mV	
5 kHz to 10 kHz	330 V to 1 000 V	0.4 % + 10 mV	



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signat Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			Fluke 8508A GIDEP/MET TRACK
10 Hz to 40 Hz	1 mV to 10 mV	0.034 % + 3 μ V	
40 Hz to 1 000 Hz	1 mV to 10 mV	0.023 % + 1.1 μ V	
1 kHz to 20 kHz	1 mV to 10 mV	0.034 % + 1.1 μ V	
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
20 kHz to 50 kHz	2 mV to 10 mV	0.12 % + 1.1 μ V	
50 kHz to 100 kHz	2 mV to 10 mV	0.57 % + 1.1 μ V	
100 kHz to 300 kHz	2 mV to 10 mV	4.6 % + 2 μ V	
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
10 Hz to 40 Hz	10 mV to 100 mV	0.01 % + 4 μ V	
40 Hz to 1 000 Hz	10 mV to 100 mV	0.01 % + 2 μ V	
1 kHz to 30 kHz	10 mV to 100 mV	0.02 % + 2 μ V	
20 kHz to 50 kHz	10 mV to 100 mV	0.034 % + 2 μ V	
50 kHz to 100 kHz	10 mV to 100 mV	0.091 % + 2 μ V	
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
100 kHz to 300 kHz	10 mV to 100 mV	0.34 % + 10 μ V	
300 kHz to 1 MHz	10 mV to 100 mV	1.2 % + 10 μ V	
1MHz to 2 MHz	10 mV to 100 mV	1.5 % + 10 μ V	
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
1 Hz to 40 Hz	0.1 V to 1 V	0.007 % + 40 μ V	
40 Hz to 1 kHz	0.1 V to 1 V	0.007 % + 20 μ V	
1 kHz to 20 kHz	0.1 V to 1 V	0.014 % + 20 μ V	
20 kHz to 50 kHz	0.1 V to 1 V	0.03 % + 20 μ V	
50 kHz to 100 kHz	0.1 V to 1 V	0.08 % + 20 μ V	
100 kHz to 300 kHz	0.1 V to 1 V	0.3 % + 100 μ V	
0.3 MHz to 1 MHz	0.1 V to 1 V	1 % + 100 μ V	
1 MHz to 2 MHz	0.1 V to 1 V	1.5 % + 100 μ V	
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
1 Hz to 40 Hz	1 V to 10 V	0.01 % + 400 μ V	
40 Hz to 1 kHz	1 V to 10 V	0.01 % + 200 μ V	
1 kHz to 20 kHz	1 V to 10 V	0.02 % + 200 μ V	
20 kHz to 50 kHz	1 V to 10 V	0.035 % + 200 μ V	
50 kHz to 100 kHz	1 V to 10 V	0.091 % + 200 μ V	
100 kHz to 300 kHz	1 V to 10 V	0.34 % + 1 mV	



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signat Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
300 kHz to 1 MHz	1 V to 10 V	1.3 % + 1 mV	Fluke 8508A GIDEP/MET TRACK
1 MHz to 2 MHz	1 V to 10 V	1.7 % + 1 mV	
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
1Hz to 40 Hz	10 V to 100 V	0.063 % + 4 mV	
40 Hz to 1 kHz	10 V to 100 V	0.035 % + 2 mV	
1 kHz to 20 kHz	10 V to 100 V	0.023 % + 2 mV	
20 kHz to 50 kHz	10 V to 100 V	0.06 % + 2 mV	
50 kHz to 100 kHz	10 V to 100 V	0.14 % + 2 mV	Fluke 8508A GIDEP/MET TRACK
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
100 kHz to 300 kHz	10 V to 100 V	0.46 % + 10 mV	
300 kHz to 1 MHz	10 V to 100 V	1.7 % + 10 mV	
Equipment to Measure AC Voltage At the list Frequencies ^{FO}			
1Hz to 40 Hz	100 V to 1 000 V	0.07 % + 40 mV	
40 Hz to 1 kHz	100 V to 1 000 V	1.9 % + 20 mV	
1 kHz to 20 kHz	100 V to 1 000 V	2.6 % + 20 mV	
20 kHz to 50 kHz	100 V to 1 000 V	0.14 % + 20 mV	
50 kHz to 100 kHz	100 V to 1 000 V	0.36 % + 20 mV	Direct measurements using Fluke 80K-40 Probe/Fluke 8508A
60 Hz	1 kV to 40 kV	0.68 %	
Equipment to measure AC Current At the listed frequencies ^{FO}			Fluke 5520A GIDEP/MET TRACK
20 kHz to 1 kHz	220 mA to 2.2 A	0.033 % + 35 μ A	
1 kHz to 5 kHz	220 mA to 2.2 A	0.053 % + 80 μ A	
5 kHz to 10 kHz	220 mA to 2.2 A	0.8 % + 0.16 mA	
Equipment to measure AC Current At the listed frequencies ^{FO}			
10 Hz to 45 Hz	1.1 A to 2.999 99 A	0.18 % + 100 μ A	
45 kHz to 1 kHz	1.1 A to 2.999 99 A	0.13 % + 100 μ A	
1 kHz to 5 kHz	1.1 A to 2.999 99 A	0.14 % + 1 000 μ A	
5 kHz to 10 kHz	1.1 A to 2.999 99 A	2 % + 5 000 μ A	Fluke 5520A GIDEP/MET TRACK
Equipment to measure AC Current At the listed Frequencies ^{FO}			
45 kHz to 100 Hz	3 A to 10.999 9 A	0.12 % + 2 mA	
0.1 kHz to 1 kHz	3 A to 10.999 9 A	0.14 % + 2 mA	
1 kHz to 5 kHz	3 A to 10.999 9 A	2.4 % + 2 mA	



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signal Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to measure AC Current At the listed Frequencies ^{FO}			Fluke 5520A GIDEP/MET TRACK
45 kHz to 100 Hz	11 A to 20.5 A	0.42 % + 5 mA	
0.1 kHz to 1 kHz	11 A to 20.5 A	0.33 % + 5 mA	
1 kHz to 5 kHz	11 A to 20.5 A	2.4 % + 5 mA	
Equipment to measure AC Current At the listed Frequencies ^{FO}			Fluke 8508A GIDEP/MET TRACK
1 kHz to 10 kHz	2 μ A to 200 μ A	0.034 % + 20 nA	
10 kHz to 10 kHz	2 μ A to 200 μ A	0.03 % + 20 nA	
10 kHz to 30 kHz	2 μ A to 200 μ A	0.03 % + 20 nA	
Equipment to measure AC Current At the listed Frequencies ^{FO}			Fluke 8508A GIDEP/MET TRACK
1 kHz to 10 kHz	200 μ A to 2 μ A	0.04 % + 0.2 μ A	
10 kHz to 10 kHz	200 μ A to 2 μ A	0.03 % + 0.2 μ A	
10 kHz to 30 kHz	200 μ A to 2 μ A	0.03 % + 0.2 μ A	
Equipment to measure AC Current At the listed Frequencies ^{FO}			Fluke 8508A GIDEP/MET TRACK
1 kHz to 10 kHz	2 mA to 20 mA	0.04 % + 2 μ A	
10 kHz to 10 kHz	2 mA to 20 mA	0.04 % + 2 μ A	
10 kHz to 30 kHz	2 mA to 20 mA	0.04 % + 2 μ A	
Equipment to measure AC Current At the listed Frequencies ^{FO}			Fluke 8508A GIDEP/MET TRACK
1 kHz to 10 kHz	20 mA to 200 mA	0.04 % + 20 μ A	
10 kHz to 10 kHz	20 mA to 200 mA	0.03 % + 20 μ A	
10 kHz to 30 kHz	20 mA to 200 mA	0.07 % + 20 μ A	
Equipment to measure AC Current At the listed Frequencies ^{FO}			Fluke 8508A GIDEP/MET TRACK
10 kHz to 2 kHz	200 mA to 2 mA	0.07 % + 0.2 mA	
2 kHz to 10 kHz	200 mA to 2 mA	0.08 % + 0.2 mA	
10 kHz to 30 kHz	200 mA to 2 mA	0.68 % + 0.2 mA	
Equipment to measure AC Current At the listed frequencies			Fluke 8508A GIDEP/MET TRACK
10 kHz to 2 kHz	2 mA to 20 mA	0.09 % + 0.2 mA	
2 kHz to 10 kHz	2 mA to 20 mA	0.57 % + 0.2 mA	
Equipment to measure AC Capacitance At the listed Frequencies ^{FO}			Fluke 5520A GIDEP/MET TRACK
10 Hz to 10 kHz	0.19 nF to 1.09 nF	0.55 % + 0.01 nF	
10 Hz to 3 kHz	1.1 nF to 3.29 nF	0.43 % + 0.01 nF	
10 Hz to 1 kHz	3.3 nF to 10.9 nF	0.19 % + 0.01 nF	
10 Hz to 1 kHz	11 nF to 109.9 nF	0.19 % + 0.1 nF	
10 Hz to 1 kHz	110 nF to 329.9 nF	0.19 % + 0.3 nF	



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signat Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to measure AC Capacitance At the listed Frequencies ^{FO}			Fluke 5520A GIDEP/MET TRACK
10 Hz to 600 Hz	0.33 μ F to 1.09 μ F	0.19 % + 1 nF	
10 Hz to 300 Hz	1.1 μ F to 3.29 μ F	0.19 % + 3 nF	
10 Hz to 150 Hz	3.29 μ F to 10.9 μ F	0.19 % + 10 nF	
10 Hz to 120 Hz	11 μ F to 32.9 μ F	0.31 % + 30 nF	
10 Hz to 80 Hz	33 μ F to 109.9 μ F	0.35 % + 100 nF	
10 Hz to 50 Hz	11 μ F to 329.9 μ F	0.35 % + 300 nF	
10 Hz to 20 Hz	0.33 μ F to 1.09 μ F	0.35 % + 1 μ F	

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Gauges ^{FO}	1 500 psi to 15 000 psi	0.1 % of reading + 0.577 psi	Crystal Xp2i GIDEP
Torque Wrenches ^{FO}	4 in·lb to 50 in·lb	0.25 % of reading + 0.18 in·lb	CDI 2000-400-02 GIDEP
	30 in·lb to 400 in·lb	0.25 % of reading + 1.39 in·lb	
	80 in·lb to 1 000 in·lb	0.25% of reading + 3.01 in·lb	
	20 ft·lb to 250 ft·lb	0.25% of reading + 0.87 ft·lb	
	60 ft·lb to 600 ft·lb	0.25% of reading + 2.08 ft·lb	CDI 2000-12-02 GIDEP

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.



Certificate of Accreditation: Supplement

The Meter Shop

6934 Signat Drive, Houston Texas, 77041
John Brown Phone: 713-957-8586

Accreditation is granted to the facility to perform the following calibrations:

4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.

